



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023

APRIL - 2023



PC-1024

Lab Name Govt. PHC, Thulukkamuthur Lab No 9270
Constituent Group Chemistry III Date of Result Entered : 19/04/2023
PT item Lyophilized human serum based Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119.7 mg/dL	0.03	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate IM	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.4 mg/dL	1.37	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109.7 mg/dL	-0.33	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Thulukkamuthur

THE MEDICAL OFFICER, GOVERNMENT HEALTH PRIMARY CENTRE, THULUKKAMUTHUR, THULUKKAMUTHUR (PO), AVINASHI (TK), TIRUPPUR (DT)

Thulukkamuthur

TAMILNADU 641654

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

ap21.wc0271
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
THULUKKAMUTHUR
AVINASHI - 641 654