



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 23.09.2023
		Calibration valid : 22.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE ,GANAPATHY, COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINNE PIPETTE f3
iii	Value	100µl-1000µl
iv	Serial No	PHC/GA/PIP/001

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.06	1000.11
2	100.03	500.05	1000.03
3	100.05	500.02	1000.07
4	100.05	500.10	1000.08
5	100.00	500.02	1000.05
6	100.08	500.07	1000.02

For Heart Beat Medical Systems,

(Handwritten signature)
(Authorized signatory)

