

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Gangalkondan  
 Lab No 12071  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[Click on the analyze to view Graphical Data](#)


[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	186 mg/dL	0.55	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	2.1 mg/dL	2.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	119 mg/dL	0.76	0.27
		<b>SDI Range</b>	<b>Interpretation</b>							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

Links

  
**MEDICAL OFFICER**  
 Govt. Primary Health Centre  
 Gangalkondan