



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt. Block PHC, Valikandapuram	Lab No	7123
Constituent Group	Chemistry III	Date of Result Entered :	21/08/2023
PT item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Year Value	SDI	II
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	172 mg/dL	-0.40	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1 mg/dL	-6.33	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Valikandapuram
The Block Medical Officer, Valikandapuram Po, Veppanthattai Tk, Perambalur Dt
Valikandapuram
TAMILNADU621115

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
GOVERNMENT PHC
VALIKANDAPURAM
TAMILNADU 621115

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Valikandapuram
 Lab No: 7123
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl. No.	Analyte	Method	Instrument	Run No.	Result	CV	Target	Unit	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	109 mg/dL	-0.06	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Pierate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.4 mg/dL	-1.23	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Home	Report History	Data Entry	Data Entry Summary
Data Entry	Data Entry	Data Entry	Data Entry

Handwritten signature and stamp of the laboratory.

June - 2023

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Valikandapuram

Details About Robust Analysis

Lab No 7123

Detail About Monthly Summary

Month June

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participant		Test value	SDI	U
						CV	SB			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	285 mg/dL	-0.11	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.1 mg/dL	-2.75	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	113 mg/dL	-0.14	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Logout

MEDICAL OFFICER
GOVERNMENT PRIMARY HEALTH CENTRE,
VALIKANDAPURAM - 621 115.

MAY 2023

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Valikandapuram
 Lab No 7123
 Month May
 Year 2023

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Mean Value	SDI	U
						CV	%SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	185 mg/dL	0.44	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	123 mg/dL	1.44	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out

MEDICAL OFFICER
 GOVERNMENT PRIMARY HEALTH CENTRE
 VALIKANDAPURAM - 621 11



CHRISTIAN MEDICAL COLLEGE
CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023

April-23



Lab Name Govt. Block PHC, Valikandapuram

Lab No 7123

Constituent Group Chemistry III

Date of Result Entered : 20/04/2023

PT item Lyophilized human serum based

Date of Report Published : 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Test Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124 mg/dL	0.63	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	112 mg/dL	0.07	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. Block PHC, Valikandapuram
 The Block Medical Officer, Valikandapuram Po, Veppanthattai Tk, Perambalur Dt
 Valikandapuram
 TAMILNADU621115

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 MEDICAL OFFICER
 GOVERNMENT PRIMARY HEALTH CENTRE,
 VALIKANDAPURAM - 621 115.



CHRISTIAN MEDICAL COLLEGE
CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name Govt. Block PHC, Valikandapuram

Lab No 7123

Constituent Group Chemistry III

Date of Result Entered : 20/03/2023

PT item Lyophilized human serum based

Date of Report Published : 04/04/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Total Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	85 mg/dL	-0.95	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	103 mg/dL	-1.37	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Valikandapuram
The Block Medical Officer, Valikandapuram Po, Veppanthattal Tk, Perambalur Dt
Valikandapuram
TAMILNADU 621115

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

[Signature]
MEDICAL OFFICER
GOVERNMENT PRIMARY HEALTH OFFICE
VALIKANDAPURAM - 621115