



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. PHC, Gangaikondan

Lab No

12071

Constituent Group Chemistry III

Date of Result Entered :

21/07/2023

PT item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	101.7 mg/dL	-1.26	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114.3 mg/dL	0.28	0.22
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Gangaikondan
 GOVT PRIMARY HEALTH CENTRE, GANGAIKONDAN
 Cuddalore
 TAMILNADU607802

Coordinator Contact Details
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

MEDICAL OFFICER
 Govt. Primary Health Centre
 Gangaikondan

[Handwritten Signature]