



CMC EXTERNAL QUALITY ASSURANCE SCHEME DEPARTMENT OF CLINICAL BIOCHEMISTRY

MONTHLY SUMMARY REPORT - APRIL 2023

Chemistry III

Lyophilized human serum based

Constituent Group

Lab Name

Govt. PHC, 4 Veerapandi

12054

Lab No

Date of Result Entered: 20/04/2023

Date of Report Published:

06/05/2023

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

CHOLESTROL III CREATININE III GLUCOSE III

CHOD-PAP III

Robonik ( Semi automatic / Fully Automatic ) Robonik ( Semi automatic / Fully Automatic ) Robonik ( Semi automatic / Fully Automatic )

> 1750 1820

111.61 2.16

Jaffes Kinetic - Alkaline Picrate III

GOD-POD III

Analyte

Method / Principle Name

Analyzer Name

Participants 1832

C

Participants

Your

119.45

6.04 8.09

7.21 0.18 5.79

116 mg/dL 2.6 mg/dL 119 mg/dL

-0.48 2.51 1.28

0.34 0.01 0.28

LAB ADDRESS:

Govt. PHC, 4 Veerapandi

THE MEDICAL OFFICER, NO 4 VEERAPANDI

TAMILNADU641019

Email:clinqc@cmcvellore.ac.in **Coordinator Contact Details:** 

Contact Number: 0416-2283102

Dr. Pamela Christudoss **CMC EQAS Coordinator** 

Panela Christudos

Christian Medical College, Vellore

CMC EQAS does not sub contract any components Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential

