

CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/23/03868

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
GANAGRISWARANPUTTAKARUNGULI-603303.

SRF No.	SBS/23/00142-0013
SRF Date	04-04-2023
Date of Receipt	04-04-2023
Date of Calibration	04-04-2023
Due Date for Calibration	03-04-2024
Issue Date	04-04-2023

Details of Unit Under Calibration

Description	Silico Pipette	Make	MICROLUX
Range	100-1000µl	Model	
Resolution	10 µl	Material	PVC
Serial Number		Operating Range	
ID Number		Condition of UUC	Good
CAL. BY	Mechanical Lab	Instrument Location	LAB

Environmental Condition

Calibration Method Used

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1030	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/06

Standard Used

Sl. No.	Description	ID.No./SI.No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP1.22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	2	3	4	5				
1	100	99.83	99.82	99.83	99.82	99.83	99.82	-0.18	0.01	0.47
		99.84	99.83	99.81	99.83	99.81				
2	500	499.96	499.97	499.96	499.97	499.96	499.94	-0.06	0.02	0.47
		499.92	499.93	499.92	499.93	499.92				
3	1000	999.86	999.87	999.86	999.86	999.87	999.90	-0.10	0.04	0.47
		999.94	999.95	999.94	999.95	999.94				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable in National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- Calibration Liquid Used: Distilled or Deionized Water (unless Graded as specified in ISO 3636)

Calibrated By,

[Signature]
(Calibration Engineer)
S.VAJINAVAN



Authorized by:

[Signature]
(Quality Manager/Chief Executive)
D.VETRI SELVI