



# Vanavil Calibrations CALIBRATION CERTIFICATE

Make It Simple, But Significant !

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/s Government Primary Health Centre Pethappampatty-6422205	Report No	VCPL/2023/002-5
	Date of Receipt	23/Mar/2023
	Date of Issue	24/Mar/2023
	Calibration Date	23/Mar/2023
	Next Due Date	22/Mar/2024
	Calibration at	On Site

### Details of Device Under Calibration:

Description:	MICRO PIPETTE	Make :	MICROLUX
Range:	5 to 50 µl	Model/Type:	-
Resolution:	1 µl	Accuracy :	As per Manual
Serial Number:	-	Location:	LAB
ID number:	GPH/PIP/01	Condition of DUC:	GOOD

Temperature	22.8 °C	National/International standard:	-
Humidity	52 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	Serial No	Certificate no	Valid Till
1	ELECTRONIC BALANCE	16031172	E150323-1	15-Mar-24

### CALIBRATION RESULT

S.NO	DUC Readings (µl)	Standard Rreadings Mean Value ( mg )	ERROR (mg)
1	10	9.96	-0.04
2	50	49.8	-0.20
3	100	99.76	-0.24

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
T.SUDHAHAR

(CALIBRATION ENGINEER)



Authorised by:

  
**S. LAKSHMANAN MURUGAN,**  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

