

b342f405-0d69-41be-ae59-d45559b67504



*12819 m medly*

External Quality Assurance Scheme - Print Monthly Summary

**CHRISTIAN MEDICAL COLLEGE**

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - AUGUST 2023



Govt. UGPHC, Kuntrathur  
 Chemistry III  
 Lyophilized human serum based

Lab No: 4595  
 Date of Result Entered: 18/08/2023  
 Date of Report Published: 01/09/2023

Sl. No.	Parameter	Method	1	2	3	4	5	6	7	8	9
1	GLUCOSE III	GOD-POD III	1841	175.58	5.13	9.00	165 mg/dL	-1.18	0.42		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01		
3	CHOLESTROL III	CHOD-PAP III	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22		

SDI Range	Interpretation
Min -1.00 to +1.00	Excellent.
Min ±1.01 to ±2.00	Good.
Min ±2.01 to ±2.99	Accept with caution. Warning Signal.
Max ±3.0	Unacceptable performance. Action Signal.

ADDRESS  
 UGPHC Kuntrathur  
 CLINICAL LABORATORY KUNDRATHUR  
 Vellore  
 TAMIL NADU-560009

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Dr. Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
 ASSISTANT SURGEON  
 UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE  
 KUNDRATHUR

Scanned with OKEN Scanner