



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name	Govt. Block PHC, Negamam	Lab No	7336
Constituent Group	Chemistry III	Date of Result Entered :	20/03/2023
PT Item	Lyophilized human serum based	Date of Report Published :	04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Mean Value	SD	CV	SI
						CV	SD				
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84	82.3 mg/dL	-1.32	2.61	
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	52.7 mg/dL	-0.39	0.98	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13	1.4 mg/dL	0.23	0.04	
4	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68	108 mg/dL	-0.62	2.20	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. Block PHC, Negamam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, COIMBATORE DISTRICT.
Negamam
TAMILNADU642120

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****

P. R. G.
22/04/23
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
NEGAMAM