

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12098

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
VALIKANDAPURAM,PERAMBALUR DISTRICT.

SRF No.	SRF/23/00478-0001
SRF Date	28-09-2023
Date of Receipt	28-09-2023
Date of Calibration	28-09-2023
Due Date for Calibration	27-09-2024
Issue Date	30-09-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	MICROLUX
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-50µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.5	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.97	9.95	9.94	9.98	9.96	9.96	-0.04	0.02	0.47
		9.93	9.92	9.95	9.98	9.97				
2	30	29.86	29.87	29.84	29.83	29.84	29.86	-0.14	0.02	0.47
		29.86	29.89	29.85	29.87	29.89				
3	50	49.95	49.97	49.93	49.95	49.98	49.90	-0.10	0.06	0.47
		49.83	49.85	49.84	49.87	49.84				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

C. Sivabalan
(Calibration Engineer)
C.SIVABALAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12099

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
VALIKANDAPURAM,PERAMBALUR DISTRICT.

SRF No.	SRF/23/00478-0002
SRF Date	28-09-2023
Date of Receipt	28-09-2023
Date of Calibration	28-09-2023
Due Date for Calibration	27-09-2024
Issue Date	30-09-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE
Resolution	1µl	Material	PVC
Serial Number	RW03038	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.7	Humidity (%RH)	56	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.85	99.85	99.85	99.85	99.95	99.89	-0.11	0.04	0.47	
		99.89	99.95	99.90	99.87	99.90					
2	500	499.90	499.95	499.90	499.90	499.95	499.90	-0.10	0.03	0.47	
		499.89	499.88	499.85	499.85	499.90					
3	1000	999.89	999.88	999.89	999.87	999.90	999.91	-0.09	0.03	0.47	
		999.93	999.90	999.91	999.94	999.95					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

C. Divahar
(Calibration Engineer)
C.SIVABALAN



Authorised by:

C. Divahar
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.