

URGENT/TIME BOUND

Heartbeat



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. Block PHC, Negamam
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No 7336
Date of Result Entered : 19/05/2023
Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Performance		CV	CV	CV
						CV	BS			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	169.6	1.04	3.43
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	31 mg/dL	0.83	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	2.1	2.01	0.05
4	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	110.8	-0.76	2.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. Block PHC, Negamam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, COIMBATORE DISTRICT,
Negamam
TAMILNADU642120

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

P. A. H.
23/05/23
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
NEGAMAM



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. Block PHC, Negamam Lab No: 7336
 Constituent Group: Chemistry III Date of Result Entered: 19/05/2023
 PT Item: Lyophilized human serum based Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Yearly Mean	CV	SD
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	169.6 mg/dL	1.04	3.43
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	31 mg/dL	0.83	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	2.1 mg/dL	2.01	0.05
4	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	110.8 mg/dL	-0.76	2.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. Block PHC, Negamam
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, COIMBATORE DISTRICT.
 Negamam
 TAMILNADU642120

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

P. B. J.
 23/09/23
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 NEGAMAM