



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. Block PHC, Negamam

Lab No 7336

Constituent Group Chemistry III

Date of Result Entered : 21/07/2023

PT Item Lyophilized human serum based

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Samples	CV	CV				
						CV	SE	CV		
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10	104.2 mg/dL	0.79	1.76
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	184	51.41	11.14	5.73	42.9 mg/dL	-1.49	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	6.9 mg/dL	3.12	0.17
4	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04	114.1 mg/dL	0.09	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Negamam  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, COIMBATORE DISTRICT.  
Negamam  
TAMILNADU642120

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not subcontract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*R. R. J.*  
23/08/23

MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
NEGAMAM