



CALIBRATION REPORT	
Report No: HBMS/CAI/22-23	Calibration date : 19.06.2023
	Calibration valid : 18.12.2023
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, NEGAMAM, COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
i Manufactured by	LABOTECH
ii Model	---
iii Serial No	769383

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

R.B.T  
 MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE  
 NEGAMAM

MD Reg No: TN/CBE/MD42/00016  
 GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
 88833 15662

No.42/1, Navaindia Road, Peelamedu,  
 CBE – 04, hbmedicalsystem@gmail.com