

THE PIPETEMANN

NABL ACCREDITED CALIBRATION LABORATORY
ADDING ACCURACY TO YOUR EQUIPMENT
MASS - VOLUME - THERMAL - SPEED - PRESSURE - TIMER - ELECTRO TECHNICAL



CALIBRATION CERTIFICATE

CUSTOMER ADDRESS : M/s. THE MEDICAL OFFICER(TPM374) GOVT PRIMARY HEALTH CENTRE VADASITHUR,KINATHUKADVU BLOCK COIMBATORE - 641202 Phone No.: 8610943045/9843245468	
UNIQUE LAB REPORT NUMBER : ULR-CC231723000004204F	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT : MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : TPMCC230959915	DATE OF ISSUE : 26-09-2023
RECEIPT NO : 12147	TEMPERATURE : 24.6 °C
RECEIPT DATE : 23-09-2023	HUMIDITY : 62 %
JOB CARD NO : TPM374092358338	AIR PRESSURE : 964_hpa
JOB CARD DATE : 23-09-2023	MEDIUM OF WEIGHING : GRADE III WATER
BRAND : FINNPIPETTE - F3	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 10 - 100 µl	DENSITY CORRECTION FACTOR [Z] : 1.0039 µl/mg
SERIAL NO : RW01521	CALIBRATION METHOD : ISO 8655:2002 PART - 6
TIPS USED : 200 µl	WEIGHING BALANCE : MICRO BALANCE
SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE (1 µl - 10000 µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT	: SARTORIUS MICRO BALANCE
SL.NO	: 28001417
RESOLUTION	: 0.000001g upto 6.1g
CALIBRATED BY	: THE PIPETEMANN(MASS),COIMBATORE
CERTIFICATE NUMBER	: TPMCCM23020421
CALIBRATION VALIDITY	: 26-02-2024
TRACEABILITY	: NATIONAL STANDARDS

Page 1 Of 2