

# CHERAN

DATE:16.06.2023

CERTIFICATE NUMBER-CBS/1014/23-24

CUSTOMER NAME : PRIMARY HEALTH CENTRE,  
SIRUMUGAI.

INSTRUMENT : MICROPIPETTE VARIABLE

MODEL : 100-1000 VARIABLE

SERIAL NUMBER : RW02650

- MICROPIPETTE 100-1000 DIS ASSEMBLED AND LUBRICATED AND FIXED.
- 100-1000 CHECKED FOR 100UL,110UL,250UL,300UL,700UL & 950UL,1000UL  
EVERYTHING IS FINE.

DATE OF CALIBRATION : 16.06.2023

NEXT CALIBRATION : 15.05.2024

THE INSTRUMENT HAS BEEN CALIBRATED SUCCESSFULLY AND FOUND TO BE WORKING  
SATISFACTORY.

FOR CHERAN BIOMEDICAL SERVICES,

AUTHORISED SIGNATORY.

SERVICE ENGINEER



  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
சிறுமுகை - 641302.  
கோவை மாவட்டம்

## CHERAN BIOMEDICAL SERVICES

PMS Building, Old 50A, New 39, Dr. Radhakrishna Road, Gandhipuram,  
Tatabad, Coimbatore - 641 012.

Ph : 0422 - 2522640, Cell : 98417 - 45417/15. Email:cheranbiomed@yahoo.co.in

GSTIN : 33ADMPR7007F1ZK, DL No. : CBE/4903/20B, CBE/4533/21B

# CHERAN

CERTIFICATE NUMBER-CBS/1013/23-24

DATE:16.06.2023

CUSTOMER NAME : PRIMARY HEALTH CENTRE,  
SIRUMUGAI.

INSTRUMENT : MICROPIPETTE VARIABLE

MODEL : 5-50 VARIABLE

SERIAL NUMBER : RW1534

- MICROPIPETTE 5-50 DIS ASSEMBLED AND LUBRICATED AND FIXED.
- 5-50 CHECKED FOR 5UL,10UL,15UL,25UL,40UL & 50UL EVERYTHING IS FINE.

DATE OF CALIBRATION : 16.06.2023

NEXT CALIBRATION : 15.05.2024

THE INSTRUMENT HAS BEEN CALIBRATED SUCCESSFULLY AND FOUND TO BE WORKING SATISFACTORY.

FOR CHERAN BIOMEDICAL SERVICES,



AUTHORISED SIGNATORY.

SERVICE ENGINEER

மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
சிறுமுகை - 641302.  
கோவை மாவட்டம்

## CHERAN BIOMEDICAL SERVICES

PMS Building, Old 50A, New 39, Dr. Radhakrishna Road, Gandhipuram,  
Tatabad, Coimbatore - 641 012.

Ph : 0422 - 2522640, Cell : 98417 - 45417/15. Email:cheranbiomed@yahoo.co.in

GSTIN : 33ADMPR7007F1ZK, DL No. : CBE/4903/20B, CBE/4533/21B