



## CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/22/14264

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  
NATTALAM, MULANKUZHI- 629162, KANYAKUMARI DISTRICT.

SRF No.	SRF/22/00323-0022
SRF Date	13-12-2022
Date of Receipt	13-12-2022
Date of Calibration	13-12-2022
Due Date for Calibration	12-12-2023
Issue Date	14-12-2022

Details of Unit Under Calibration

Description	Micro Pipette	Make	LABOPETTE
Range	1000µl	Model	
Resolution	FIXED	Material	PVC
Serial Number		Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LAB

Environmental Condition

Calibration Method Used

Temperature (°C)	23.2	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	1000	1000.68	1000.67	1000.66	1000.65	1000.66	1000.66	0.66	0.01	0.47
		1000.66	1000.66	1000.67	1000.66	1000.67				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)  
M.BALAJI



Authorised by:

(Quality Manager/Chief Executive)  
D.VETRI SELVI

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.