

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Gnanagiriswaranpetal

[Details About Robust Analysis](#)

Lab No 12214

[Detail About Monthly Summary](#)

Month March

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	0.82	96 mg/dL	0.65	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	8.41	0.13	1.5 mg/dL	1.01	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.06	6.65	120 mg/dL	1.18	2.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
--------------------------------	--------------------------------	------------------------------------	--

<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>
--------------------------------	---------------------------------	----------------------	--------------------------