

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Gnanagiriswranpetal

Details About Robust Analysis

Lab No 12214

Detail About Monthly Summary

Month February

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 21/02/2023

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robosik ( Semi automatic / Fully Automatic )	1868	155.11	6.56	10.18	153 mg/dL	-0.21	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	37	3.35	7.70	0.26	3.3 mg/dL	-0.19	0.08
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	113.39	6.52	7.39	125 mg/dL	1.57	2.57

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>