

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Granagitswranpetal

[Details About Robust Analysis](#)

Lab No 12214

[Detail About Monthly Summary](#)

Month April

[Detail about SDI](#)

Year 2023

Constituent  
Group Chemistry III[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1032	119.45	6.04	7.21	125 mg/dL	0.77	0.34
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2.1 mg/dL	-0.55	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	112 mg/dL	-0.12	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within +2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>