

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Gnanagirikwarpatal	Details About Robust Analysis								
Lab No	12214	Detail About Monthly Summary								
Month	May	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data										
<a href="#">All Analyser Result</a> <a href="#">Print</a>										
Date of Result Entered : 20/05/2023										
Date of Report Published : 05/06/2023										
S.No.	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	177 mg/dL	-0.41	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	36	1.78	8.91	0.16	1.9 mg/dL	0.75	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	110 mg/dL	-0.87	2.49
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within +2.01 to +2.99		Accept with caution. Warning Signal.								
Beyond +3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>