



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt. PHC, Gnanagiriswaranpetal	Lab No	12214
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lysophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle / Reagent	Analyser Name	No of Participants	CV	Participant		Your Value	SDI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robnik (Semi automatic / Fully Automatic)	1041	175.58	5.13	0.00	180 mg/dL	0.48	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.75	6.47	0.11	1.75 mg/dL	0.00	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.95	3.90	4.49	115 mg/dL	0.00	1.50

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance, Action Signal.

LAB ADDRESS :

Govt. PHC, Gnanagiriswaranpetal
THE MEDICAL OFFICER, GSPETTAI
Chengalpattu
TAMILNADU

Coordinator Contact Details:
Email:clinacc@cmcvellore.ac.in
Contact Number: 0416-2263102

Pamela Christudasa
Dr. Pamela Christudasa
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****