



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. Block PHC, Medavakkam Lab No 3512
 Constituent Group Chemistry III Date of Result Entered : 19/08/2023
 PT item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	173 mg/dL	-0.29	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :

Govt. Block PHC, Medavakkam

"The Chief Civil Surgeon Govt Upgraded PHC Mambakkam Main road, Medavakkam

Medavakkam

TAMILNADU600100

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

[Handwritten Signature]
28/9/23

BLOCK MEDICAL OFFICER

UP.Gr. Govt. Primary Health Centre