

**VIEW LAB MONTHLY SUMMARY**

**Lab Name** Govt. PHC, Suralkayur **Lab No** 12824 **Details of Graphic Data**  
**Month & Year** August 2023 **Sample No** 8 - 01/08/2023 **Back**  
**Constituent Group** Chemistry III **Analyte** CHOLESTROL III **CLICK HERE TO VIEW HISTOGRAM**

**METHOD : CHOD-PAP III**

*Methodwise - Analyzers Results data*

|  | Total Lab   | DV           | CV         | SD          |
|--|-------------|--------------|------------|-------------|
| <b>All Analyzers</b>                         | <b>1963</b> | <b>113.5</b> | <b>3.9</b> | <b>4.41</b> |
| Biosystems                                   | 33          | 115          | 3.9        | 4.49        |
| Transasia / Erba                             | 17          | 111.4        | 3.2        | 3.54        |
| Robonik ( Semi automatic / Fully Automatic ) | 1764        | 113.8        | 4.1        | 4.84        |
| Any Analyser (Automation / Semi Automation ) | 149         | 114          | 4.4        | 4.97        |

▲ Your Result 110.6 mg/dL **SDI** -0.69  
**RMSDI** -0.29  
 Mean for Comparison 113.8

*Levey Jennings Chart (SDI)*

