

AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33AOSPA9153Q1ZA * DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Vadugapatti, Kumarapalayam Panjayath-638106.	Certificate No:	AMS/LAB/289/4903
		Issuing Date:	26/09/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	VPHC/LAB/VMP/001
Make:	Microlux	Department:	Laboratory
Model:	---	Calibrated On:	25/09/2023
Sl. No:	---	Calibration Due Suggested:	24/09/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	5 to 50µl		

Traceability Details:

SL.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-01	31/01/2024

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	05	5.0	5.0	+0.0
2	10	10.0	10.0	+0.0
3	20	20.0	20.0	+0.0
4	30	30.0	30.0	+0.0
5	50	50.0	50.0	+0.0

Allowable Tolerance: User Defined



Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)