

SJV Scientific Company

20, Railway Lane Street, 1st Cross, Chekkan Thottam, Ganapathy, Coimbatore - 641 006.
E-mail: svj.scientific@gmail.com. Ph no: 97866-90789
GSTIN: 33AJXPR6161K1ZG

Ref -2023/Cal/PHC/SP/017/01

To,
The Medical Officer,
Govt Primary Health Centre,
Somayampalayam,
Coimbatore.

CALIBRATION REPORT

REMI CENTRIFUGE MACHINE CALIBRATION.

- Fitted with resiliently mounted motor for vibration free performance.
- A specially wound 5 step transformer is used for speed regulation.
- The highly sophisticated elegant body of the equipment is made with fiber re-enforced plastic which gives protection against corrosion and electric shock.

Technical Data:

Maximum Speed :4000 RPM,
Maximum RCF:1880 x g,
Motor: 1/20 HP AC/DC universal.
Serial No: C852

Comments: This above-mentioned product is working as expected.

Calibration Date : 27.09.2023

Due Date : 26.09.2024

SJV Scientific Company

For SJV Scientific Company

V. K. R.
Proprietor

Sales & Service:
Blood Glucometer - BP Apparatus - Weighing Scale - Nebulizer - ECG Machine - Lab Kits & Chemicals - Surgical Instruments.