

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. PHC, Maraimalai Nagar**

Lab No 10178

Month **August**

Year **2023**

Constituent Group **Chemistry III**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	Participants SD	Your Value	CV	Participants SD	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	185 mg/dL	1.05	0.42	
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	32.6 mg/dL	1.00	0.65	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01	
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	118 mg/dL	0.91	0.22	
Interpretation											
Within -1.00 to +1.00		Excellent.									
Within +1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Signature
MEDICAL OFFICER

GOVT PRIMARY HEALTH CENTRE
 MARAIMALAI NAGAR-603 203,
 CHENGALPATTU DISTRICT.