

VIEW LAB MONTHLY SUMMARY


Lab Name	Govt. PHC, Adi Annamalai	Details About Robust Analysis	
Lab No	12944	Detail About Monthly Summary	
Month	July	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		
Click on the analyte to view Graphical Data		<a href="#">All Analyser Result</a>	<a href="#">Print</a>
Date of Result Entered : 20/07/2023			
Date of Report Published : 07/08/2023			

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	115 mg/dL	0.93	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

  
 மருத்துவ அலுவலர்,  
 அரசு ஆரம்ப சுகாதார நிலையம்,  
 அடி அண்ணாமலை-606 604