


VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Adi Annamalai				Details About Robust Analysis					
Lab No	12944				Detail About Monthly Summary					
Month	May				Detail about SDI					
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data					<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
					Date of Result Entered : 20/05/2023					
					Date of Report Published : 05/06/2023					
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	182 mg/dL	0.12	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
 மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 அடி-அண்ணாமலை-606 604