



<b>CALIBRATION REPORT</b>	
Report No : HBMS/CAL/23-24	Calibration Date : 12.09.2023 Calibration valid : 11.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE , TISAIYANVILAI, THIRUNELVELI</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
<b>i</b> Manufactured by	LABOTECH
<b>ii</b> Model	BD-153
<b>iii</b> Serial No	PHC/TV/CENT/001

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1501	1
4	2000	2004	4

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com