



**REPORT ON
ELECTRICAL SAFETY TESTING/ PERFORMANCE ANALYSIS/ CALIBRATION**

Report No : TR/APHC/022/23-24

Page 1 Of 2

Calibration Date : 08/03/2023

Calibration Due : 07/03/2024

1.1 CUSTOMER DETAILS

Name and address of the organisation

Primary Health Centre
New Gh Road, Opposite State Bank Of India, Ayalur,
Gobichettipalayam, Erode District,
Tamil Nadu. 638453
Letter dated 08-03-2023
08/03/2023

Reference and Date

Date of receipt of item

1.2 DESCRIPTION OF DEVICE UNDER TEST (DUT)

NOMENCLATURE

Manufactured by

Model

Serial No.

Biomedical Product ID

Supply

Device Type

Device Classification

Location

BIO CHEMISTRY ANALYZER
ROBONIK
PRIESTEST TOUCH
AT1110211RBK
APHC/LAB/BCA/01
220V-240V AC
Type CF
Class II equipment
LABORATORY

1.3 CONDITION OF THE ITEM WHEN RECEIVED

No visible damage and in working order

1.4 ENVIRONMENTAL CONDITION OF MEASUREMENTS

A. Temperature

28.6°C

B. Relative Humidity

45-75%

C. Ambient Barometric Pressure

756mmHg

1.5 Applicable Specification

IEC Specification IEC 60601-1, IEC 60601 -2-27

1.6 Test Done

Electrical Safety and Performance Testing

Tested by :

Approved by :

Balamuralikrishnan K
(Biomedical Engineer)

Priya M
(Quality Manager)



FL BIOMEDICAL
189, Vasantham Paradise,
Chithode, Erode-638102.
Cell: 7092848995



1.7 STATUS

Manuf Specification	Users Specification	Within Specification	Out of Specification	Calibration	Electrical Safety Test	Performance Analysis
✓	-	✓	-	✓	✓	✓

1.8 TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING

SI No	Name of the Instrument	Make	Model	Serial No	Cal Due	Traceability Reference
2	ESA	RGM	288+	05H-0600	14-Aug-23	Annexure 3

2.0 PERFORMANCE ANALYSIS OF BIO CHEMISTRY ANALYZER

ELECTRICAL SAFETY TEST

SI.No	Parameter	Observed value	Cal & M.capl	Acceptable limits as per the Std.	Remarks
1	Protective earth resistance	0.23 Ω	±3%	0.0 Ω - 0.3 Ω	ok
2	Chassis Leakage	27μA	±5%	1 μA - 100 μA NC	ok
		68 μA	±5%	1 μA - 500 μA SFC	ok
3	Patient leakage current	262 μA	±5%	1 μA - 100 μA B & BF(NC)	ok
		10 μA	±5%	1 μA - 10 μA CF (SFC)	ok
4	Earth Leakage	915 μA	±5%	1 μA - 1000 μA B.BF.CF (NC)	ok
		1868 μA	±5%	1 μA - 5000 μA B.BF.CF (SFC)	ok
5	Insulation Resistance 500 V DC	Success	±5%	≥2 MΩ	ok
6	Equipment current	0.1 A	-	As per manufacture spec	ok
7	Mains voltage	233V	-	As per manufacture spec	ok

3.0 REMARKS

3.1 This report is applicable to the sample tested only.

3.2 The instruments used for testing are under valid calibration and are traceable to National Standards.

3.3 Parameter of the DUT were verified and found to be within the specified limits.

3.4 refer NABL Doc No. 121 Clause 7.0 Accommodation and environmental Conditions sub Clause see 7.2.11 below in line with ISO/IEC 17025:2005 Clause 5.3

Tested by :

Balamuralikrishnan K
(Biomedical Engineer)



Approved by :

Priya M
(Quality Manager)

FL BIOMEDICAL
189, Vasantham Paradise,
Chithode, Erode-638102,
Cell: 7092848995