

ROOTS METROLOGY & TESTING LABORATORY

(A Division of Roots Industries India Limited)

156, Bommasandra Industrial Area, Bommasandra, Jigani Link Road, Anekal Taluk, Bangalore - 560 099, Mobile : 97900 14066, E-mail : blrmetlab@rootsemail.com



CALIBRATION CERTIFICATE

ADDING VALUE		
		ULR-CC278222030011004F
		: 21-10-2022
	Date of Issue	: 18-10-2022
Certificate No. : BRMTL/01/522100770-A80	Date of Receipt of Item	: 18-10-2022
Name and address of the Customer	Date of Calibration	: 17-10-2023
M/s Medayakkam Govt. Upgraded Primary Health	Calibration Due	
Center 4th Street, Mandavali, Green court,	(As Per Customer Request)	: On-Site # 18-10-2022
Medavakkam, Chennai, Tamilnadu 600 100.	Customer Reference	: On-Site # 10 15
1. DESCRIPTION OF DEVICE UNDER CALIBRATION	(DUC)	C : Satisfactory
1. DESCRIPTION OF DEVICE UNDER CALIBRATION ELECTRICAL SAFETY (CEL	L Receiving Condition of DUC	
Name of the DUC : ELECTRICAL SALES COUNTER)	Decument	: As Per Manufacturer

Medavakkam, Chemian		ON (DUC) ON (DUC) ON (DUC)	: Satisfactory
Name of the DUC	: ELECTRICAL SAFETY (COUNTER) : UPHC/LAB/CC/01	CELL Receiving Condition of DUC Reference Document	: As Per Manufacturer Specification
Identification Sl.No. Make	: NA : SYSMAX	Cal. Procedure Ref. No. Environmental Condition	: BRMTL/WI/418 : Temp: 24.2°C / RH: 57%
Model Location	: XP-100 : LABORATORY	Calibrated at	: Onsite ole To / Certificate Validity

2. DETAILS OF MASTER U	ISED		Traceable To / Certificate	Validity
	ng Isa / Model	SL. No. / ID No.	No.	12.09.2023
Standard Instrument Used	Na Market Co.	48K-0413	ACCS221794	12.09.2023
Electrical Safety Analyzer	Rigel Medical / 288+	BRMTL-315 48K-0413	CRMTL/01/422100328-A1	22.04.2023
	Rigel Medical / 288+	BRMTL-315	A	
Electrical Safety Analyzer	1 / International standard	ds as per ISO/IEC 1702	25	

The Masters are traceable to National / International standards as per ISO/IEC 17025

FIELD - MEDICAL DEVICES

Calibration Results

	Remarks	
. VISUAL INSPECTION Description	Pass Fail	
L. No.	1	
Device Clean and Decontaminated	V	
2 Physical damage	V	
Power cord check Display intensity adequate for day time use	V .	
5 Alarm Function		
EX ECTRICAL SAFETY		

4. ELECTRICAL SAFETY

4. ELECT	TRICAL SATE TO	OLTAGE)	Allowable Limit (V)	Uncertainty ± (%)		
1	Specification	Macaured Value (V)	230 ± 10 %	7.59	Pass Pass	
1	Live to Neutral	231	230 ± 10 %	7.37		

CALIBRATED BY (CALIBRATION ENGINEER)		i and and may not be reproduced in full	
		or in part, without written permission as Quality Manager Roots Metrology & Toeting Laboratory, Bangalore.	
DINESHKUMAR A	KAMALAKANNAN R		

* NEXT CALIBRATION DUE AS PER CUSTOMER'S PLAN AND REQUIREMENT