



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Alangayam

Lab No

3042

Constituent Group Chemistry III

Date of Result Entered :

18/03/2023

PT Item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl No	Analyte	Method	Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	Z
							CY	SD			
1	GLUCOSE III	GOD-POD III		Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	89 mg/dL	-0.36	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III		Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	108 mg/dL	-0.50	0.28
SDI Range					Interpretation						
Within -1.00 to +1.00					Excellent.						
Within ±1.01 to ±2.00					Good.						
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.						
Beyond ±3.0					Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Alangayam
MEDICAL OFFICER ALANGAYAM, VELLORE DISTRICT
Alangayam
TAMILNADU635701

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

A. S. Pan
MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
ALANGAYAM, 635 701

Homogeneity and Stability of the sample is passed.

Data In CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****