



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. Block PHC, Arisipalayam
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7334
Date of Result Entered: 19/08/2023
Date of Report Published: 01/09/2023

| SL.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 177 mg/dL | 0.16 | 0.42 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.7 mg/dL | -0.50 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 114 mg/dL | 0.04 | 0.22 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. Block PHC, Arisipalayam
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, COIMBATORE DISTRICT.
Arisipalayam
TAMILNADU641032

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

[Signature]
Medical Officer
Govt. Primary Health C
Arisipalayam
Coimbatore 32