



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. Block PHC, Alangayam

Lab No

3042

Constituent Group Chemistry III

Date of Result Entered :

28/04/2023

PT Item Lyophilized human serum based

Date of Report Published :

06/05/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	126 mg/dL	0.91	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109 mg/dL	-0.45	0.28
		<b>SDI Range</b>	<b>Interpretation</b>							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

**LAB ADDRESS :**

Govt. Block PHC, Alangayam  
MEDICAL OFFICER ALANGAYAM, VELLORE DISTRICT  
Alangayam  
TAMILNADU635701

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christodoss*  
Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

**MEDICAL OFFICER**  
**COMMUNITY HEALTH CENTRE**  
**ALANGAYAM - 635 028**

*G. S. J. Per*

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*