



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name	Govt. Block PHC, Alangayam	Lab No	3042
Constituent Group	Chemistry III	Date of Result Entered :	19/05/2023
PT item	Lyophilized human serum based	Date of Report Published :	05/06/2023

Sl. No.	Parameter	Method	Target	Observed	Delta	Delta Z	Delta X	Delta Y	Delta Z	Delta X	Delta Y
1	GLUCOSE III	GOD-POD III	1834	180.88	5.19	9.39	190 mg/dL	0.97	0.44		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1854	1.76	7.48	0.13	2 mg/dL	1.82	0.01		
3	CHOLESTROL III	CHOD-PAP III	1798	114.60	5.08	5.83	110 mg/dL	-0.79	0.27		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Alangayam
 MEDICAL OFFICER ALANGAYAM, VELLORE DISTRICT
 Alangayam
 TAMILNADU635701

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Govt PHC
 COMMUNITY HEALTH CENTER
 ALANGAYAM - 635701

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****