



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name Govt. Block PHC, Alangayam

Lab No

3042

Constituent Group Chemistry III

Date of Result Entered

19/06/2023

PT Item Lyophilized human serum based

Date of Report Published

04/07/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	270 mg/dL	-1.12	0.89
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.1 mg/dL	-2.75	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24
		<b>SDI Range</b>	<b>Interpretation</b>							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

**LAB ADDRESS :**

Govt. Block PHC, Alangayam  
MEDICAL OFFICER ALANGAYAM, VELLORE DISTRICT  
Alangayam  
TAMILNADU635701

Coordinator Contact Details:  
Email: clinq@cmcvellore.ac.in  
Contact Number: 0415-2283102

*Pamela Christodoss*  
Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

1. COMMUNITY HEALTH CENTRE  
ALANGAYAM - 635 701

*G. S. P. S.*

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*