




CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. Block PHC, Alangayam

Lab No

3042

Constituent Group Chemistry III

Date of Result Entered :

19/07/2023

PT Item Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method	Principle Name	Analyzer Name	No of Participants	DV	Participate CV	SD	Target Value	>DI	L
1	GLUCOSE III	GOD-POD III		Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
3	CHOLESTROL III	CHOD-PAP III		Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	104 mg/dL	-1.93	0.22
SDI Range					Interpretation						
Within -1.00 to +1.00					Excellent						
Within ±1.01 to ±2.00					Good.						
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.						
Beyond ±3.0					Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Alangayam
MEDICAL OFFICER ALANGAYAM, VELLORE DISTRICT
Alangayam
TAMILNADU635701

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER

COMMUNITY HEALTH CENTRE
ALANGAYAM . 635 701

Homogeneity and Stability of the sample is passed.

Data In CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****