



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. UGPHC, Muthupettai
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 7269
 Date of Result Entered: 20/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	192 mg/dL	1.82	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	27 mg/dL	-0.30	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	105 mg/dL	-1.90	0.22
Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. UGPHC, Muthupettai
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVARUR DISTRICT.
 Muthupettai
 TAMILNADU 614704

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

MEDICAL OFFICER
 Govt. Primary Health Centre
 MUTUPPETI - 614 704
 Thiruvavur District