

CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date: 23.09.2023
	Calibration Valid: 22.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S. CORPORATION HOSPITAL, UPHC,THIDIYALUR,COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	ERBA
ii Model	ERBAPETTE
iii Value	100µl-1000µl
iv Serial No	01161181


Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.06	1000.11
2	100.03	500.05	1000.03
3	100.05	500.02	1000.07
4	100.09	500.10	1000.08
5	100.00	499.98	1000.09
6	100.08	500.07	1000.02

For Heart Beat Medical Systems,


(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com

CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 23.09.2023
	Calibration Valid: 22.03.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/S. CORPORATION HOSPITAL, UPHC, THUDIYALUR, COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	P FACT
II Model	5µl - 50µl
III Serial No	392286
IV Location	LABORATORY

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.01	20.03	50.05
2	05.03	20.04	50.02
3	05.01	20.05	50.04
4	05.03	20.02	50.02
5	05.01	20.06	50.05
6	05.02	20.02	50.04

For Heart Beat Medical Systems,

(Handwritten signature)
(Authorized signatory)

