



CHRISTIAN MEDICAL COLLEGE.

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. PHC, Indur
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 12113
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic	1841	175.58	5.13	9.00	225 mg/dL	5.49	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic	1764	113.80	4.07	4.64	100 mg/dL	-2.98	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Indur
 THE MEDICAL OFFICER, INDUR PHC, NALLAMPALLI, DHARMAPURI DIST.
 Indur
 TAMILNADU636803

Coordinator Contact Details:
 Email: clnqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

MEDICAL OFFICER
GPHC
 INDUR (PO), NALLAMPALLI (TK)
 DHARMAPURI (DT) - 636 803.

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****