



VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name Govt. UPHC, Savariyarpalayam  
 Lab No 10719  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

[Back to Search Page](#)[Print](#)

Date of Result Entered : 18/08/2023

Date of Report Published : 01/09/2023

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48			3.25
			Transasia / Erba	23	173.17	4.83	8.36			3.49
			Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	191 mg/dL	1.71	0.42
			Any Analyser (Automation / Semi Automation )	64	174.34	5.46	9.51			2.38
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	59	28.92	13.99	4.05	47.2 mg/dL	4.51	1.05
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11			0.04
			Transasia / Erba	21	1.74	6.34	0.11			0.05
			Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2.6 mg/dL	7	0.01
			Any Analyser (Automation / Semi Automation )	62	1.76	7.18	0.13			0.03
4	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49			1.56
			Transasia / Erba	17	111.42	3.18	3.54			1.72
			Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22
			Any Analyser (Automation / Semi Automation )	149	113.96	4.36	4.97			0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out