

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. PHC, Mekkalur

Lab No: 12925

Participant Category: Chemistry III

Date of Result Entered: 20/06/2023

Participant: Lyophilized human serum based

Date of Report Published: 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	276 mg/dL	-0.72	0.69
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.2 mg/dL	-2.16	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116 mg/dL	0.45	0.24

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
±1.01 to ±2.00	Good.
±2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :
 Govt. PHC, Mekkalur
 Mekkalur
 Vellore, TAMIL NADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not subcontract any components
 ***** End of Report *****