

3d3ad302-0e37-42fb-a9d6-d7914552d14e



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023



4595

Lab No

18/05/2023

Date of Result Entered :

05/06/2023

Date of Report Published :

Govt. UGPHC, Kunrathur

Chemistry III

Lyophilized human serum based

Parameter	Method	1834	180.88	5.19	9.39	196 mg/dL	1.61	0.44	
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	143 mg/dL	4.87	0.27
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)							

SDI Range	Interpretation
SDI ≤ 1.00	Excellent.
SDI ≤ 1.01 to ≤ 2.00	Good.
SDI ≤ 2.01 to ≤ 2.99	Accept with caution. Warning Signal.
SDI > ≤ 3.0	Unacceptable performance. Action Signal.

ADDRESS :
 Govt. UGPHC, Kunrathur
 CLINICAL LABORATORY KUNDRATHUR
 Kunrathur
 TAMILNADU 600069

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

✓
 மேம்பட்ட சர்டிபிகேட்
 அரசு ஆரம்ப சுகாதார நிலையம்
 குன்றத்தூர், சென்னை - 600069

Scanned with OKEN Scanner



Lock screen



Play



Share



Search

