

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Papparasatti  
 Lab No: 12128  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robonik Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Results](#) [Print](#)

Date of Result Entered : 13/08/2023


Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Used	Analysed By	No of Participants	CV	Target	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.50	185.25 mg/dL	1.09	0.42	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	6.12	1.58 mg/dL	-6.87	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1754	112.80	4.87	4.64	114.75 mg/dL	0.20	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Monthly Summary</a>	<a href="#">Monthly Methods</a>	<a href="#">Send Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Daily Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Logout</a>

  
**MEDICAL OFFICER**  
 Govt Primary Health Centre  
 PAPPARAPATTY - 635 809  
 Dharmapuri Dt.