



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 15.09.2023
		Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE , JAMUNAMARATHUR, THIRUVANNAMALAI.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	REMI
ii	Model	ZHBN
iii	Serial No	05090

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	2001	1

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range


Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non – contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :




MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com