

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE , KRISHNANKOVIL-629 001		Certificate No:	VCPL/CC/23-24/22-001
		Date of Calibration:	29-09-2023
		Date of issue:	30-09-2023
		Date of Receipt:	29-09-2023
		Next cal Date:	28-09-2024
Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	-
Range:	(5 to 50) μ l	Model/Type:	-
Resolution:	0.2 μ l	Condition of DUC	GOOD
Serial Number	235847	Location:	LABORATORY
ID number	-	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	23.4 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	48 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	5.0	5.02	-0.02
2	25.0	24.97	0.03
3	50.0	49.96	0.04

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH

(CALIBRATION ENGINEER)



Authorised by:


T. SUDHAHAR,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyar Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE , KRISHNANKOVIL - 629 001.		Certificate No: VCPL/CC/23-24/22-002
		Date of Calibration: 29-09-2023
		Date of issue: 30-09-2023
		Date of Receipt: 29-09-2023
		Next cal Date: 28-09-2024
Details of Device Under Calibration		
Description:	MICROPIPETTE	Make: DRAGON LAB
Range:	(100 to 1000) μ l	Model/Type: -
Resolution:	1 μ l	Condition of DUC GOOD
Serial Number	YE16BAA0097203	Location: LABORATORY
ID number	-	Cal. At: ONSITE
Environmental Conditions		Calibration Method Used
Temperature	23.2 °C	Calibration Method Used ISO 8655:2002 PART-6
Humidity	49 % Rh	Cal procedure no: -

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTech	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100	99.92	0.08
2	500	499.88	0.12
3	1000	999.84	0.16


Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH

(CALIBRATION ENGINEER)



Authorised by:


T. SUDHAKAR,
Technical Manager

Vanavil Calibrations Pvt L^t

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE, KRISHNANKOVIL - 629 001.	Certificate No:	VCPL/CC/23-24/22-003
	Date of Calibration:	29-09-2023
	Date of issue:	30-09-2023
	Date of Receipt:	29-09-2023
	Next cal Date:	28-09-2024

Details of Device Under Calibration

Description:	MICROSCOPE	Make:	LYNX
Range:	-	Model/Type:	ASZ-N107T
Resolution:	-	Condition of DUC	GOOD
Serial Number	008473	Location:	LABORATORY
ID number	-	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	26.2 °C	National/International standard:	IEC60601-1
Humidity	51 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24

Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Leakage	166.3	<5000 μ A For B,BF,CF
		MEASURED VALUES μA	ALLOWABLE LIMITS μA
2	Enclosure Leakage	202.3	<5000 μ A For B,BF,CF

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH

(CALIBRATION ENGINEER)



Authorised by:


T. SUDHAKAR,
Technical Manager

Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

52, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



Vanavil Calibrations

Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page : 1 of 1

Customer Name & Address: M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE , KRISHNANKOVIL-629 001.	Certificate No:	VCPL/CC/23-24/22-004
	Date of Calibration:	29-09-2023
	Date of issue:	30-09-2023
	Date of Receipt:	29-09-2023
	Next cal Date:	28-09-2024

Details of Device Under Calibration

Description	CENTRIFUGE	Make:	REMI
Range	3500 RPM	Model/Type:	C-854/8
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	ZBIN-27100	Instrument Location:	LABORATORY

Environmental Conditions

Temperature	25.4 °C	Reference Standard No	SANAS TR4501
Humidity	53 % Rh	Calibration Procedure No	-

Method Used For Calibration

Details of Reference Standard Equipment For Calibration

S.No	Description	ID No / SL No	Certificate No	Validity
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005 SYSTEMS/ HTM 560	4-Dec-23


Calibration Result

S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	500	500.2	-0.2	12
2	1000	999.7	0.3	12
3	2000	1998	2	12
4	3000	2997	3	12
5	3500	3503	-3	12

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:


C.AKASH
(CALIBRATION ENGINEER)



Authorised by:

T. SUDHAKAR,
Technical Manager
Vanavil Calibrations Pvt l

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranal, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com