

8/9/23, 3:07 PM

External Quality Assurance Scheme - View Monthly Summary

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Amathur	Details About Robust Analysis
Lab No	13146	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	111 mg/dL	0.27	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.6 mg/dL	-0.66	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Early Summary	Change Password	Home	Sign-out

N. Rajasekar
மருத்துவ அலுவலர்
ஆமத்தூர் அரசு ஆரம்ப
சுகாதார நிலையம்
விருதுநகர் - 5.