



Trust & Care



CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 26.09.2023
		Calibration Valid: 25.03.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, MEKKALURE, THIRUVANNAMALAI.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	ERBA
II	Model	---
III	Value	10µl
IV	Serial No	PHC/PIP/005

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	10µl
1	10.05
2	10.06
3	10.01
4	09.97
5	10.01
6	10.02

For Heart Beat Medical Systems



(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com