



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Melaramanathi

Lab No: 12559

Constituent Group: Chemistry III

Date of Result Entered: 19/04/2023

PT Item: Lyophilized human serum based

Date of Report Published: 06/05/2023

Sl. No.	Test Name	Method	Result	Reference Range	Units	CV (%)	CV (%)	CV (%)	CV (%)
1	GLUCOSE III	GOD-POD III	74	118.33	5.77	6.83	115 mg/dL	-0.49	1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	65	2.14	7.02	0.15	2.2 mg/dL	0.40	0.04
3	CHOLESTROL III	CHOD-PAP III	143	111.22	4.31	4.79	120 mg/dL	1.83	0.80
SDI Range		Interpretation							
Within -1.00 to +1.00		Excellent.							
Within ±1.01 to ±2.00		Good.							
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.							
Beyond ±3.0		Unacceptable performance. Action Signal.							

**LAB ADDRESS :**

Govt. PHC, Melaramanathi  
 Melaramanathi  
 TAMILNADU

**Coordinator Contact Details:**  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential